RADIATION PROTECTION SECTION



Division of Health Service Regulation • N.C. Department of Health and Human Services

SERVICE COMPANY RADIATION PROTECTION PROGRAM GUIDE

Each registrant shall develop, document, and implement a Radiation Protection Program commensurate with the scope and extent of registered activities to ensure compliance with the provisions of **10A NCAC 15** [.1600]. In accordance with Rule [.1603 (c)], the registrant shall annually review the radiation protection program content and implementation.

N.C. Radiation Dictionary

N.C. Radiation Regulations

REQUIREMENTS TO INCLUDE IN RADIATION PROTECTION PROGRAM

Radiation Protection Program: [.1603(a)]

ALARA (As Low As Reasonably Achievable) Radiation Protection Principles

• Describe procedures and engineering controls used, based upon sound radiation principles to achieve occupational doses, and doses to members of the public ALARA. [.1603(b)]

Personnel

- Describe how individuals are notified if occupational doses exceed 1 mSv (100 mrem) TEDE or 1 mSv (100 mrem) to any individual organ or tissue. **[.1004]** (effective date 1/1/14)
- State the company's annual occupational dose limits. [.1604(a)] Dose Limits
- Explain how doses that an individual may receive in the current year are reduced by the occupational exposures received while employed by another person. [.1604(f)]
- State the dose limits to an Embryo/Fetus? [.1610] [.1614(1)(c)]
- Explain the company's personnel voluntary declared pregnancy policy. [.1640(f)] Pregnancy Policy
- Describe the company's personnel monitoring exposure policy. [.1614]
 - If personnel monitoring not provided to operators, explain how the company met compliance to the regulations.
- Describe the company's process to obtain prior occupational dose for new workers. [.1638(a)(1)&(2)]
- Explain the company's retention period for exposure records. (keep all dosimetry reports)
 [.1640(a)(1)&(g)]

Exceeding Exposure Limits [.1647)] Dose Limits

- Identify the company's reporting process for when a dose limit is exceeded? [.1646][.1647]
- Describe how data of the affected person is reported to both the individual and to Radiation Protection.
 - [.1647(b)(c)(d)(e)] [.0111] - Estimated dose
 - Cause of elevated exposure
 - Corrective Action
 - Name & Date of Birth
 - The Last 4 of the Social security number and/or employee identifier

REVIEW OF RADIATION PROTECTION PROGRAM

In accordance with Rule .1603 (c), the registrant shall annually review the radiation protection program content and implementation.

Signature	Date	

The N.C. Radiation link http://www.ncradiation.net/Xray/reference.htm has the following information:		
<u>Inspection Check List</u>	Information the inspectors ask to see during an inspection	
<u>Tips For Answering Violations</u>	How to reply to violations	
Reference Guides For Facilities	Information to help the facilities write their own safety program	
Plan Reviews & Surveys	Registered service providers that are qualified to provide this service	
Registration Requirements	Requirements for new and existing facilities	
Steps To Install X-ray Equipment or Register	Guide for installing equipment	
Registration Forms	Form to register a facility/X-ray machines or amend a registration	
Registration: Owning or Possessing X-ray Equipment	Guide for Owning or Possessing X-ray Equipment in N.C.	
<u>Postings</u>	Documents that need to be posted in the facility	

Links of Interest:

AAPM Reports	http://www.aapm.org/pubs/reports/
AAPM Medical Physics Practice	http://www.aapm.org/pubs/MPPG/
Guidelines	
CT Protocols	http://www.aapm.org/pubs/CTProtocols/